

## MRI Safety Information

Parameter	Notes
Product name	<b>aerstent TBJ</b>
Manufacturer	bess pro gmbh Gustav-Krone-Str. 7 D—14167 Berlin Germany
Distributor	Leufen Medical GmbH Gustav-Krone-Str. 7 D—14167 Berlin Germany Tel.: +49 30 816 90 93 00 Fax: +49 30 816 90 93 93 <a href="http://www.leufen-medical.eu">www.leufen-medical.eu</a> <a href="mailto:contact@leufen-medical.eu">contact@leufen-medical.eu</a>
Document number and edition date	MRITBJ- 2 — 2023-01
Location of MRI safety information	<a href="http://www.leufen-medical.eu/tbj">www.leufen-medical.eu/tbj</a>
Static magnetic field strength	1.5 T / 3 T
Type of nuclei	Hydrogen
BO Field Orientation	Horizontal
Maximum Spatial Field Gradient	19 T/m
Maximum B0*IdB0/drl product	48 T <sup>2</sup> /m
RF Transmit Coil Type	Whole Body or extremities transmit RF coils (trunk and head local transmit coils are excluded)
RF Receive Coil Type	Any receive RF coil can be used.
RF Power Conditions	Normal Operating Mode
Maximum Whole-Body SAR	Whole body SAR ≤ 2 W/kg ATTENTION: Under the scan conditions defined above, the product <i>aerstent TBJ</i> is expected to produce the following maximum temperature rise after 15 minutes of continuous scanning: <ul style="list-style-type: none"> <li>• 1.5 T: 4.1 ± 0.5 °C</li> <li>• 3 T: 4.0 ± 0.5 °C</li> </ul> Tested REF: 533-20-050
Patient Landmarking Criteria	No restrictions
Patient Position in Scanner	No restrictions
MR Image Artifact	The presence of a product of the type <i>aerstent TBJ</i> may produce an image artifact. Some manipulation of scan parameters may be required to compensate for the artifact. In vitro assessment on a comparable product ( <i>aixstent OES</i> ) at 1.5T showed: <ul style="list-style-type: none"> <li>• On spin echo sequences a maximal artifact width of 25 mm (surrounding the device)</li> <li>• On the corresponding gradient echo sequences a maximal artifact width of 15 mm (surrounding the device)</li> </ul>